TRU-2000Refraction Unit

Display your clinical excellence confidently with the TOMEY refraction unit TRU-2000. Its high functionality, combined with classic design and the comfortable chair, provide a reliable basis for all eye examinations.

Excellent features

- + Space-saving
- + Preconfigured for TOMEY equipment
- + Personal touch with internal LED light
- + Easy to install (right or left side)
- + Timeless yet innovative
- + Comfortable motorised chair with tilting action



Specifications

| FEATURES | |
|----------------------------|---|
| Table | Rotating 90°, laterally movable, for 2 devices |
| Table height adjustment | Not available |
| Table lock | Magnetic |
| Max. load capacity (table) | 50 kg |
| Patient chair | ER-1000 |
| Patient chair elevation | 200 mm |
| Max. load capacity (chair) | 150 kg |
| Projector column height | approx. 1700 mm |
| Connectors | 2x power outlet (230 V) 1× RS-232, 1× 6-pin (6 V / 12 V), 1× LAN, 1× USB 2.0 |
| Reading light | White LED 12 V |
| Internal light | RGB LED 12 V |

| ELECTRIC REQUIREMENTS & WEIGHT | |
|--------------------------------|-----------------------------|
| Power consumption | 280 VA |
| Power supply | 110/115 V AC/230 V AC ±10 % |
| Total weight (incl. chair) | approx. 270 kg |
| | |

| DIMENSIONS | |
|------------|-----------------------|
| Width | max. 1440 mm (±10 mm) |
| Height | max. 1750 mm (±10 mm) |
| Depth | max. 2360 mm (±10 mm) |
| Table | 870 x 440 mm (±5 mm) |

| OPTIONAL | |
|--------------------|---|
| | PA-1000 Phoropter arm "balance" |
| | PA-3000 Phoropter arm "linear" |
| Optional equipment | SH-1000 Skiascope/Ophthalmoscope holder |
| | TC-1000 Trial lens cabinet |
| | CBS-1000 Control box stand |



TOMEY EUROPE TOMEY GMBH

tomey.de

TOMEY GmbH is the European headquarter of TOMEY Corporation, 2-11-33 Noritakeshinmachi Nishi-Ku, Nagoya, 451-0051, Japan

2023/08 – subject to change without notice

Always read and follow the instructions for use

Not all products, services or offers are approved or offered in every market. Please note that the current status of approval for the labelling, instructions and contents of the brochure may vary from one country to another.

